

Client ID: RAW NORTH (IN001)

Lab ID: 2601079-01

CAS #	Analyte	Result	RL	Units	Dilution	Analyzed Date	QC Batch	Method	Analyst	Qualifier
Organics-PFAS Isotope Dilution										
763051-92-9	11Cl-PF3OUdS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
356-02-5	3:3FTCA	ND	4.0	ng/L	1	01/22/26	B6A1420	8327	BM	
757124-72-4	4:2FTS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
914637-49-3	5:3FTCA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
27619-97-2	6:2FTS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
812-70-4	7:3FTCA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
39108-34-4	8:2FTS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
756426-58-1	9Cl-PF3ONS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
919005-14-4	ADONA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
13252-13-6	HFPO-DA	ND	4.0	ng/L	1	01/22/26	B6A1420	8327	BM	
2991-50-6	NEtFOSAA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
2355-31-9	NMeFOSAA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
375-22-4	PFBA	1.8	4.0	ng/L	1	01/22/26	B6A1420	8327	BM	J
375-73-5	PFBS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
30334-69-1	PFBSA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
335-76-2	PFDA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
307-55-1	PFDoDA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
335-77-3	PFDS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
646-83-3	PFECHS	0.8	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	J
375-85-9	PFHpA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
375-92-8	PFHpS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
307-24-4	PFHxA	1.3	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	J
355-46-4	PFHxS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
41997-13-1	PFHxSA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
375-95-1	PFNA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
68259-12-1	PFNS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
335-67-1	PFOA	1.8	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	J
1763-23-1	PFOS	1.6	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	J
754-91-6	PFOSA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
2706-90-3	PFPeA	1.3	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	J
2706-91-4	PFPeS	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
376-06-7	PFTeDA	ND	4.0	ng/L	1	01/22/26	B6A1420	8327	BM	
72629-94-8	PFTTrDA	ND	4.0	ng/L	1	01/22/26	B6A1420	8327	BM	
2058-94-8	PFUnDA	ND	2.0	ng/L	1	01/22/26	B6A1420	8327	BM	
<i>Isotope Dilution Analog: 13C2-4:2FTS</i>			115 %		47-213	01/22/26	B6A1420	8327	BM	
<i>Isotope Dilution Analog: 13C2-6:2FTS</i>			104 %		10-169	01/22/26	B6A1420	8327	BM	
<i>Isotope Dilution Analog: 13C2-8:2FTS</i>			101 %		15-176	01/22/26	B6A1420	8327	BM	
<i>Isotope Dilution Analog: 13C2-PFDoDA</i>			97.2 %		51-139	01/22/26	B6A1420	8327	BM	
<i>Isotope Dilution Analog: 13C2-PFDoDA+13C2-PFTeDA</i>			85.2 %		38-128	01/22/26	B6A1420	8327	BM	
<i>Isotope Dilution Analog: 13C2-PFTeDA</i>			63.6 %		10-123	01/22/26	B6A1420	8327	BM	
<i>Isotope Dilution Analog: 13C3-HFPODA</i>			110 %		67-127	01/22/26	B6A1420	8327	BM	

Client ID: RAW NORTH (IN001)

Lab ID: 2601079-01

CAS #	Analyte	Result	RL	Units	Dilution	Analyzed Date	QC Batch	Method	Analyst	Qualifier
Organics-PFAS Isotope Dilution										
	<i>Isotope Dilution Analog: 13C3-PFBS</i>		102 %		67-115	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C3-PFHxS</i>		103 %		87-116	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C4-PFBA</i>		106 %		72-121	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C4-PFHpA</i>		111 %		87-117	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C5-PFHxA</i>		111 %		85-116	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C5-PFPeA</i>		105 %		83-116	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C6-PFDA</i>		106 %		72-123	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C7-PFUnDA</i>		101 %		73-121	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C8-PFOA</i>		117 %		82-117	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C8-PFOS</i>		103 %		84-114	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C8-PFOA</i>		104 %		84-117	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: 13C9-PFNA</i>		108 %		83-123	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: d3-N-MeFOSAA</i>		111 %		76-130	01/22/26	B6A1420	8327	BM	
	<i>Isotope Dilution Analog: d5-N-EtFOSAA</i>		104 %		74-136	01/22/26	B6A1420	8327	BM	

Organics-PFAS Isotope Dilution - Quality Control

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Analyzed	Qualifier
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	----------	-----------

Batch B6A1420 - Method: 3512

Prepared: 01/14/2026

Blank (B6A1420-BLK1)

11Cl-PF3OUdS	ND	ng/L							01/21/2026	
3:3FTCA	ND	ng/L							01/21/2026	
4:2FTS	ND	ng/L							01/21/2026	
5:3FTCA	ND	ng/L							01/21/2026	
6:2FTS	ND	ng/L							01/21/2026	
7:3FTCA	ND	ng/L							01/21/2026	
8:2FTS	ND	ng/L							01/21/2026	
9Cl-PF3ONS	ND	ng/L							01/21/2026	
ADONA	ND	ng/L							01/21/2026	
HFPO-DA	1.12	ng/L							01/21/2026	J
NEtFOSAA	ND	ng/L							01/21/2026	
NMeFOSAA	ND	ng/L							01/21/2026	
PFBA	ND	ng/L							01/21/2026	
PFBS	ND	ng/L							01/21/2026	
PFBSA	ND	ng/L							01/21/2026	
PFDA	ND	ng/L							01/21/2026	
PFDoDA	ND	ng/L							01/21/2026	
PFDS	ND	ng/L							01/21/2026	
PFECHS	ND	ng/L							01/21/2026	
PFHpA	ND	ng/L							01/21/2026	
PFHpS	ND	ng/L							01/21/2026	
PFHxA	ND	ng/L							01/21/2026	
PFHxS	ND	ng/L							01/21/2026	
PFHxSA	ND	ng/L							01/21/2026	
PFNA	ND	ng/L							01/21/2026	
PFNS	ND	ng/L							01/21/2026	
PFOA	ND	ng/L							01/21/2026	
PFOS	ND	ng/L							01/21/2026	
PFOSA	ND	ng/L							01/21/2026	
PFPeA	ND	ng/L							01/21/2026	
PFPeS	ND	ng/L							01/21/2026	
PFTeDA	ND	ng/L							01/21/2026	
PFTTrDA	ND	ng/L							01/21/2026	
PFUnDA	ND	ng/L							01/21/2026	
Isotope Dilution Analog: 13C2-4:2FTS	150	ng/L	150.1		99.7	47-213			01/21/2026	
Isotope Dilution Analog: 13C2-6:2FTS	139	ng/L	152.2		91.5	10-169			01/21/2026	
Isotope Dilution Analog: 13C2-8:2FTS	151	ng/L	153.6		98.6	15-176			01/21/2026	
Isotope Dilution Analog: 13C2-PFDoDA	150	ng/L	160.0		93.8	51-139			01/21/2026	
Isotope Dilution Analog: 13C2-PFDoDA+13C2-PFTeDA	152	ng/L	160.0		95.0	38-128			01/21/2026	
Isotope Dilution Analog: 13C2-PFTeDA	156	ng/L	160.0		97.2	10-123			01/21/2026	
Isotope Dilution Analog: 13C3-HFPODA	318	ng/L	320.0		99.4	67-127			01/21/2026	
Isotope Dilution Analog: 13C3-PFBS	149	ng/L	149.1		100	67-115			01/21/2026	
Isotope Dilution Analog: 13C3-PFHxS	158	ng/L	151.7		104	87-116			01/21/2026	
Isotope Dilution Analog: 13C4-PFBA	167	ng/L	160.0		104	72-121			01/21/2026	
Isotope Dilution Analog: 13C4-PFHpA	161	ng/L	160.0		101	87-117			01/21/2026	
Isotope Dilution Analog: 13C5-PFHxA	161	ng/L	160.0		101	85-116			01/21/2026	
Isotope Dilution Analog: 13C5-PFPeA	165	ng/L	160.0		103	83-116			01/21/2026	
Isotope Dilution Analog: 13C6-PFDA	160	ng/L	160.0		100	72-123			01/21/2026	
Isotope Dilution Analog: 13C7-PFUnDA	162	ng/L	160.0		102	73-121			01/21/2026	
Isotope Dilution Analog: 13C8-PFOA	168	ng/L	160.0		105	82-117			01/21/2026	

Organics-PFAS Isotope Dilution - Quality Control

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Analyzed	Qualifier
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	----------	-----------

Batch B6A1420 - Method: 3512

Prepared: 01/14/2026

Blank (B6A1420-BLK1)

<i>Isotope Dilution Analog: 13C8-PFOS</i>	155	ng/L	153.4		101	84-114			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	166	ng/L	160.0		104	84-117			01/21/2026	
<i>Isotope Dilution Analog: 13C9-PFNA</i>	166	ng/L	160.0		104	83-123			01/21/2026	
<i>Isotope Dilution Analog: d3-N-MeFOSAA</i>	161	ng/L	160.0		100	76-130			01/21/2026	
<i>Isotope Dilution Analog: d5-N-EtFOSAA</i>	163	ng/L	160.0		102	74-136			01/21/2026	

LCS (B6A1420-BS1)

11Cl-PF3OUds	39.6	ng/L	37.80		105	76-132			01/21/2026	
3:3FTCA	47.1	ng/L	40.00		118	68-138			01/21/2026	
4:2FTS	38.6	ng/L	37.40		103	79-132			01/21/2026	
5:3FTCA	43.3	ng/L	40.00		108	75-130			01/21/2026	
6:2FTS	38.0	ng/L	38.00		100	80-130			01/21/2026	
7:3FTCA	43.0	ng/L	40.00		108	76-131			01/21/2026	
8:2FTS	41.7	ng/L	38.40		109	79-133			01/21/2026	
9Cl-PF3ONS	39.7	ng/L	37.40		106	77-133			01/21/2026	
ADONA	40.8	ng/L	37.80		108	79-133			01/21/2026	
HFPO-DA	41.1	ng/L	40.00		103	73-138			01/21/2026	
NEtFOSAA	39.6	ng/L	40.00		99.0	78-131			01/21/2026	
NMeFOSAA	44.2	ng/L	40.00		110	79-130			01/21/2026	
PFBA	42.5	ng/L	40.00		106	77-131			01/21/2026	
PFBS	37.1	ng/L	35.40		105	82-130			01/21/2026	
PFBSA	42.2	ng/L	40.00		106	78-125			01/21/2026	
PFDA	44.7	ng/L	40.00		112	78-132			01/21/2026	
PFDoDA	45.0	ng/L	40.00		113	80-132			01/21/2026	
PFDS	40.3	ng/L	38.60		104	76-134			01/21/2026	
PFECHS	37.7	ng/L	36.96		102	75-136			01/21/2026	
PFHpA	41.9	ng/L	40.00		105	78-131			01/21/2026	
PFHpS	41.7	ng/L	38.20		109	76-132			01/21/2026	
PFHxA	42.8	ng/L	40.00		107	79-135			01/21/2026	
PFHxS	38.1	ng/L	36.50		105	78-132			01/21/2026	
PFHxSA	42.6	ng/L	40.00		107	79-125			01/21/2026	
PFNA	46.7	ng/L	40.00		117	80-128			01/21/2026	
PFNS	38.7	ng/L	38.40		101	79-133			01/21/2026	
PFOA	43.9	ng/L	40.00		110	78-132			01/21/2026	
PFOS	39.4	ng/L	37.04		106	80-133			01/21/2026	
PFOSA	42.2	ng/L	40.00		106	80-127			01/21/2026	
PFPeA	43.5	ng/L	40.00		109	76-135			01/21/2026	
PFPeS	38.8	ng/L	37.60		103	76-131			01/21/2026	
PFTeDA	46.2	ng/L	40.00		115	80-130			01/21/2026	
PFTTrDA	42.2	ng/L	40.00		105	74-127			01/21/2026	
PFUnDA	39.6	ng/L	40.00		99.0	80-133			01/21/2026	
<i>Isotope Dilution Analog: 13C2-4:2FTS</i>	162	ng/L	150.1		108	47-213			01/21/2026	
<i>Isotope Dilution Analog: 13C2-6:2FTS</i>	140	ng/L	152.2		92.2	10-169			01/21/2026	
<i>Isotope Dilution Analog: 13C2-8:2FTS</i>	159	ng/L	153.6		104	15-176			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFDoDA</i>	154	ng/L	160.0		96.1	51-139			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFDoDA+13C2-PFTeDA</i>	155	ng/L	160.0		96.7	38-128			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFTeDA</i>	157	ng/L	160.0		97.8	10-123			01/21/2026	
<i>Isotope Dilution Analog: 13C3-HFPODA</i>	327	ng/L	320.0		102	67-127			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFBS</i>	146	ng/L	149.1		98.0	67-115			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFHxS</i>	157	ng/L	151.7		103	87-116			01/21/2026	

Organics-PFAS Isotope Dilution - Quality Control

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Analyzed	Qualifier
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	----------	-----------

Batch B6A1420 - Method: 3512

Prepared: 01/14/2026

LCS (B6A1420-BS1)

<i>Isotope Dilution Analog: 13C4-PFBA</i>	168	ng/L	160.0		105	72-121			01/21/2026	
<i>Isotope Dilution Analog: 13C4-PFHpA</i>	164	ng/L	160.0		102	87-117			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFHxA</i>	163	ng/L	160.0		102	85-116			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFPeA</i>	160	ng/L	160.0		100	83-116			01/21/2026	
<i>Isotope Dilution Analog: 13C6-PFDA</i>	167	ng/L	160.0		105	72-123			01/21/2026	
<i>Isotope Dilution Analog: 13C7-PFUnDA</i>	164	ng/L	160.0		102	73-121			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	164	ng/L	160.0		103	82-117			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOS</i>	157	ng/L	153.4		102	84-114			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	165	ng/L	160.0		103	84-117			01/21/2026	
<i>Isotope Dilution Analog: 13C9-PFNA</i>	160	ng/L	160.0		100	83-123			01/21/2026	
<i>Isotope Dilution Analog: d3-N-MeFOSAA</i>	163	ng/L	160.0		102	76-130			01/21/2026	
<i>Isotope Dilution Analog: d5-N-EtFOSAA</i>	163	ng/L	160.0		102	74-136			01/21/2026	

LCS Dup (B6A1420-BSD1)

11Cl-PF3OUdS	39.2	ng/L	37.80		104	76-132	1.07	14	01/21/2026	
3:3FTCA	44.0	ng/L	40.00		110	68-138	6.76	18.6	01/21/2026	
4:2FTS	40.3	ng/L	37.40		108	79-132	4.36	13.8	01/21/2026	
5:3FTCA	45.5	ng/L	40.00		114	75-130	4.96	13.5	01/21/2026	
6:2FTS	39.8	ng/L	38.00		105	80-130	4.68	13.3	01/21/2026	
7:3FTCA	39.4	ng/L	40.00		98.5	76-131	8.83	13.7	01/21/2026	
8:2FTS	40.8	ng/L	38.40		106	79-133	2.23	16.7	01/21/2026	
9Cl-PF3ONS	38.9	ng/L	37.40		104	77-133	1.98	12.8	01/21/2026	
ADONA	39.9	ng/L	37.80		105	79-133	2.33	12.5	01/21/2026	
HFPO-DA	46.2	ng/L	40.00		115	73-138	11.5	19.8	01/21/2026	
NEtFOSAA	42.0	ng/L	40.00		105	78-131	5.89	14.1	01/21/2026	
NMeFOSAA	42.6	ng/L	40.00		107	79-130	3.59	13.5	01/21/2026	
PFBA	42.4	ng/L	40.00		106	77-131	0.0942	11.8	01/21/2026	
PFBS	36.2	ng/L	35.40		102	82-130	2.56	13	01/21/2026	
PFBSA	42.2	ng/L	40.00		106	78-125	0.0947	13.5	01/21/2026	
PFDA	44.5	ng/L	40.00		111	78-132	0.538	12.5	01/21/2026	
PFDODA	44.9	ng/L	40.00		112	80-132	0.356	12.5	01/21/2026	
PFDS	39.1	ng/L	38.60		101	76-134	2.97	15	01/21/2026	
PFECHS	37.9	ng/L	36.96		102	75-136	0.476	14	01/21/2026	
PFHpA	41.5	ng/L	40.00		104	78-131	1.10	12.5	01/21/2026	
PFHpS	42.5	ng/L	38.20		111	76-132	1.85	15.4	01/21/2026	
PFHxA	43.4	ng/L	40.00		109	79-135	1.58	12.6	01/21/2026	
PFHxS	39.0	ng/L	36.50		107	78-132	2.28	14.7	01/21/2026	
PFHxSA	41.7	ng/L	40.00		104	79-125	2.33	12.8	01/21/2026	
PFNA	41.7	ng/L	40.00		104	80-128	11.4	13.3	01/21/2026	
PFNS	38.8	ng/L	38.40		101	79-133	0.155	14	01/21/2026	
PFOA	42.4	ng/L	40.00		106	78-132	3.57	13	01/21/2026	
PFOS	39.2	ng/L	37.04		106	80-133	0.458	12.4	01/21/2026	
PFOSA	41.6	ng/L	40.00		104	80-127	1.58	12	01/21/2026	
PFPeA	42.9	ng/L	40.00		107	76-135	1.39	11.4	01/21/2026	
PFPeS	40.2	ng/L	37.60		107	76-131	3.54	14.5	01/21/2026	
PFTeDA	44.1	ng/L	40.00		110	80-130	4.61	13	01/21/2026	
PFTrDA	43.4	ng/L	40.00		108	74-127	2.76	15.7	01/21/2026	
PFUnDA	42.6	ng/L	40.00		106	80-133	7.30	13.6	01/21/2026	
<i>Isotope Dilution Analog: 13C2-4:2FTS</i>	158	ng/L	150.1		105	47-213			01/21/2026	
<i>Isotope Dilution Analog: 13C2-6:2FTS</i>	104	ng/L	152.2		68.4	10-169			01/21/2026	
<i>Isotope Dilution Analog: 13C2-8:2FTS</i>	156	ng/L	153.6		102	15-176			01/21/2026	

Organics-PFAS Isotope Dilution - Quality Control

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Analyzed	Qualifier
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	----------	-----------

Batch B6A1420 - Method: 3512

Prepared: 01/14/2026

LCS Dup (B6A1420-BSD1)

<i>Isotope Dilution Analog: 13C2-PFDoDA</i>	160	ng/L	160.0		100	51-139			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFDoDA+13C2-PFTeDA</i>	165	ng/L	160.0		103	38-128			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFTeDA</i>	175	ng/L	160.0		109	10-123			01/21/2026	
<i>Isotope Dilution Analog: 13C3-HFPODA</i>	320	ng/L	320.0		99.9	67-127			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFBS</i>	146	ng/L	149.1		97.6	67-115			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFHxS</i>	150	ng/L	151.7		98.8	87-116			01/21/2026	
<i>Isotope Dilution Analog: 13C4-PFBA</i>	167	ng/L	160.0		105	72-121			01/21/2026	
<i>Isotope Dilution Analog: 13C4-PFHpA</i>	167	ng/L	160.0		104	87-117			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFHxA</i>	167	ng/L	160.0		104	85-116			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFPeA</i>	164	ng/L	160.0		103	83-116			01/21/2026	
<i>Isotope Dilution Analog: 13C6-PFDA</i>	167	ng/L	160.0		105	72-123			01/21/2026	
<i>Isotope Dilution Analog: 13C7-PFUnDA</i>	165	ng/L	160.0		103	73-121			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	167	ng/L	160.0		104	82-117			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOS</i>	157	ng/L	153.4		103	84-114			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	165	ng/L	160.0		103	84-117			01/21/2026	
<i>Isotope Dilution Analog: 13C9-PFNA</i>	173	ng/L	160.0		108	83-123			01/21/2026	
<i>Isotope Dilution Analog: d3-N-MeFOSAA</i>	162	ng/L	160.0		101	76-130			01/21/2026	
<i>Isotope Dilution Analog: d5-N-EtFOSAA</i>	161	ng/L	160.0		100	74-136			01/21/2026	

MRL Check (B6A1420-MRL1)

11Cl-PF3OUdS	3.50	ng/L	3.780		92.6	76-131			01/21/2026	
3:3FTCA	4.88	ng/L	4.000		122	62-156			01/21/2026	
4:2FTS	3.42	ng/L	3.740		91.4	73-139			01/21/2026	
5:3FTCA	4.26	ng/L	4.000		106	72-139			01/21/2026	
6:2FTS	3.78	ng/L	3.800		99.5	65-144			01/21/2026	
7:3FTCA	4.32	ng/L	4.000		108	72-139			01/21/2026	
8:2FTS	4.36	ng/L	3.840		114	68-140			01/21/2026	
9Cl-PF3ONS	3.40	ng/L	3.740		90.9	78-128			01/21/2026	
ADONA	3.58	ng/L	3.780		94.7	78-129			01/21/2026	
HFPO-DA	5.70	ng/L	4.000		142	53-155			01/21/2026	
NEtFOSAA	3.80	ng/L	4.000		95.0	68-143			01/21/2026	
NMeFOSAA	4.02	ng/L	4.000		100	70-133			01/21/2026	
PFBA	4.10	ng/L	4.000		102	64-139			01/21/2026	
PFBS	3.32	ng/L	3.540		93.8	70-134			01/21/2026	
PFBSA	4.32	ng/L	4.000		108	78-131			01/21/2026	
PFDA	5.00	ng/L	4.000		125	75-132			01/21/2026	
PFDoDA	4.24	ng/L	4.000		106	79-127			01/21/2026	
PFDS	3.82	ng/L	3.860		99.0	71-140			01/21/2026	
PFECHS	3.12	ng/L	3.696		84.4	74-139			01/21/2026	
PFHpA	3.96	ng/L	4.000		99.0	73-132			01/21/2026	
PFHpS	3.60	ng/L	3.820		94.2	66-144			01/21/2026	
PFHxA	4.42	ng/L	4.000		110	76-133			01/21/2026	
PFHxS	3.92	ng/L	3.650		107	77-137			01/21/2026	
PFHxSA	4.16	ng/L	4.000		104	78-130			01/21/2026	
PFNA	3.84	ng/L	4.000		96.0	67-140			01/21/2026	
PFNS	3.52	ng/L	3.840		91.7	74-137			01/21/2026	
PFOA	4.52	ng/L	4.000		113	64-141			01/21/2026	
PFOS	3.50	ng/L	3.704		94.5	76-144			01/21/2026	
PFOSA	4.34	ng/L	4.000		108	78-132			01/21/2026	
PFPeA	3.68	ng/L	4.000		92.0	74-136			01/21/2026	
PFPeS	3.74	ng/L	3.760		99.5	74-135			01/21/2026	

Organics-PFAS Isotope Dilution - Quality Control

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Analyzed	Qualifier
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	----------	-----------

Batch B6A1420 - Method: 3512

Prepared: 01/14/2026

MRL Check (B6A1420-MRL1)

PFTeDA	4.48	ng/L	4.000		112	76-130			01/21/2026	
PFTrDA	4.48	ng/L	4.000		112	71-124			01/21/2026	
PFUnDA	3.76	ng/L	4.000		94.0	73-133			01/21/2026	
<i>Isotope Dilution Analog: 13C2-4:2FTS</i>	161	ng/L	150.1		107	47-213			01/21/2026	
<i>Isotope Dilution Analog: 13C2-6:2FTS</i>	107	ng/L	152.2		70.2	10-169			01/21/2026	
<i>Isotope Dilution Analog: 13C2-8:2FTS</i>	161	ng/L	153.6		105	15-176			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFDoDA</i>	153	ng/L	160.0		95.8	51-139			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFTeDA</i>	151	ng/L	160.0		94.4	38-128			01/21/2026	
<i>13C2-PFDoDA+13C2-PFTeDA</i>										
<i>Isotope Dilution Analog: 13C2-PFTeDA</i>	147	ng/L	160.0		92.1	10-123			01/21/2026	
<i>Isotope Dilution Analog: 13C3-HFPODA</i>	325	ng/L	320.0		101	67-127			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFBS</i>	146	ng/L	149.1		98.2	67-115			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFHxS</i>	157	ng/L	151.7		103	87-116			01/21/2026	
<i>Isotope Dilution Analog: 13C4-PFBA</i>	167	ng/L	160.0		104	72-121			01/21/2026	
<i>Isotope Dilution Analog: 13C4-PFHpA</i>	163	ng/L	160.0		102	87-117			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFHxA</i>	161	ng/L	160.0		101	85-116			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFPeA</i>	162	ng/L	160.0		102	83-116			01/21/2026	
<i>Isotope Dilution Analog: 13C6-PFDA</i>	160	ng/L	160.0		100	72-123			01/21/2026	
<i>Isotope Dilution Analog: 13C7-PFUnDA</i>	156	ng/L	160.0		97.3	73-121			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	164	ng/L	160.0		103	82-117			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOS</i>	162	ng/L	153.4		105	84-114			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOXA</i>	162	ng/L	160.0		101	84-117			01/21/2026	
<i>Isotope Dilution Analog: 13C9-PFNA</i>	165	ng/L	160.0		103	83-123			01/21/2026	
<i>Isotope Dilution Analog: d3-N-MeFOSAA</i>	169	ng/L	160.0		106	76-130			01/21/2026	
<i>Isotope Dilution Analog: d5-N-EtFOSAA</i>	168	ng/L	160.0		105	74-136			01/21/2026	

Matrix Spike (B6A1420-MS1)

Source: 2601050-01

11Cl-PF3OUdS	39.4	ng/L	37.80	ND	104	71-129			01/21/2026	
3:3FTCA	43.7	ng/L	40.00	ND	109	70-138			01/21/2026	
4:2FTS	38.9	ng/L	37.40	ND	104	78-135			01/21/2026	
5:3FTCA	47.9	ng/L	40.00	ND	120	75-131			01/21/2026	
6:2FTS	38.5	ng/L	38.00	ND	101	75-137			01/21/2026	
7:3FTCA	41.3	ng/L	40.00	ND	103	73-134			01/21/2026	
8:2FTS	38.5	ng/L	38.40	ND	100	76-135			01/21/2026	
9Cl-PF3ONS	40.3	ng/L	37.40	ND	108	76-136			01/21/2026	
ADONA	41.8	ng/L	37.80	ND	111	79-139			01/21/2026	
HFPO-DA	40.8	ng/L	40.00	ND	102	71-141			01/21/2026	
NEtFOSAA	42.2	ng/L	40.00	ND	106	77-135			01/21/2026	
NMeFOSAA	41.1	ng/L	40.00	ND	103	77-133			01/21/2026	
PFBA	43.9	ng/L	40.00	1.44	106	70-141			01/21/2026	
PFBS	38.2	ng/L	35.40	ND	108	79-137			01/21/2026	
PFBSA	42.6	ng/L	40.00	0.380	106	75-133			01/21/2026	
PFDA	44.9	ng/L	40.00	ND	112	77-135			01/21/2026	
PFDoDA	44.7	ng/L	40.00	ND	112	79-134			01/21/2026	
PFDS	38.2	ng/L	38.60	ND	98.9	68-132			01/21/2026	
PFECHS	39.8	ng/L	36.96	ND	108	62-153			01/21/2026	
PFHpA	42.2	ng/L	40.00	ND	105	77-134			01/21/2026	
PFHpS	39.6	ng/L	38.20	ND	104	75-134			01/21/2026	
PFHxA	43.6	ng/L	40.00	1.26	106	74-142			01/21/2026	
PFHxS	37.4	ng/L	36.50	ND	103	72-141			01/21/2026	
PFHxSA	42.2	ng/L	40.00	ND	106	76-133			01/21/2026	

Organics-PFAS Isotope Dilution - Quality Control

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Analyzed	Qualifier
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	----------	-----------

Batch B6A1420 - Method: 3512

Prepared: 01/14/2026

Matrix Spike (B6A1420-MS1) Source: 2601050-01

PFNA	43.5	ng/L	40.00	ND	109	77-134			01/21/2026	
PFNS	40.2	ng/L	38.40	ND	105	76-134			01/21/2026	
PFOA	41.9	ng/L	40.00	1.18	102	71-141			01/21/2026	
PFOS	41.6	ng/L	37.04	1.08	110	71-144			01/21/2026	
PFOSA	41.3	ng/L	40.00	ND	103	78-131			01/21/2026	
PFPeA	42.9	ng/L	40.00	ND	107	72-142			01/21/2026	
PFPeS	38.8	ng/L	37.60	ND	103	70-139			01/21/2026	
PFTeDA	44.9	ng/L	40.00	ND	112	78-133			01/21/2026	
PFTrDA	43.4	ng/L	40.00	ND	108	66-136			01/21/2026	
PFUnDA	40.8	ng/L	40.00	ND	102	79-135			01/21/2026	
<i>Isotope Dilution Analog: 13C2-4:2FTS</i>										
	156	ng/L	150.1		104	47-213			01/21/2026	
<i>Isotope Dilution Analog: 13C2-6:2FTS</i>										
	171	ng/L	152.2		112	10-169			01/21/2026	
<i>Isotope Dilution Analog: 13C2-8:2FTS</i>										
	161	ng/L	153.6		105	15-176			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFDoDA</i>										
	157	ng/L	160.0		98.3	51-139			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFDoDA+13C2-PFTeDA</i>										
	149	ng/L	160.0		93.3	38-128			01/21/2026	
<i>Isotope Dilution Analog: 13C2-PFTeDA</i>										
	135	ng/L	160.0		84.4	10-123			01/21/2026	
<i>Isotope Dilution Analog: 13C3-HFPODA</i>										
	333	ng/L	320.0		104	67-127			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFBS</i>										
	144	ng/L	149.1		96.9	67-115			01/21/2026	
<i>Isotope Dilution Analog: 13C3-PFHxS</i>										
	155	ng/L	151.7		102	87-116			01/21/2026	
<i>Isotope Dilution Analog: 13C4-PFBA</i>										
	163	ng/L	160.0		102	72-121			01/21/2026	
<i>Isotope Dilution Analog: 13C4-PFHpA</i>										
	164	ng/L	160.0		103	87-117			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFHxA</i>										
	171	ng/L	160.0		107	85-116			01/21/2026	
<i>Isotope Dilution Analog: 13C5-PFPeA</i>										
	162	ng/L	160.0		101	83-116			01/21/2026	
<i>Isotope Dilution Analog: 13C6-PFDA</i>										
	169	ng/L	160.0		106	72-123			01/21/2026	
<i>Isotope Dilution Analog: 13C7-PFUnDA</i>										
	166	ng/L	160.0		104	73-121			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>										
	181	ng/L	160.0		113	82-117			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOS</i>										
	152	ng/L	153.4		98.7	84-114			01/21/2026	
<i>Isotope Dilution Analog: 13C8-PFOA</i>										
	164	ng/L	160.0		102	84-117			01/21/2026	
<i>Isotope Dilution Analog: 13C9-PFNA</i>										
	173	ng/L	160.0		108	83-123			01/21/2026	
<i>Isotope Dilution Analog: d3-N-MeFOSAA</i>										
	169	ng/L	160.0		106	76-130			01/21/2026	
<i>Isotope Dilution Analog: d5-N-EtFOSAA</i>										
	166	ng/L	160.0		104	74-136			01/21/2026	

Matrix Spike Dup (B6A1420-MSD1) Source: 2601050-01

11Cl-PF3OUds	37.9	ng/L	37.80	ND	100	71-129	3.83	25.7	01/21/2026	
3:3FTCA	44.5	ng/L	40.00	ND	111	70-138	1.90	26.7	01/21/2026	
4:2FTS	38.5	ng/L	37.40	ND	103	78-135	1.03	25.1	01/21/2026	
5:3FTCA	45.3	ng/L	40.00	ND	113	75-131	5.54	26	01/21/2026	
6:2FTS	39.6	ng/L	38.00	ND	104	75-137	2.87	24	01/21/2026	
7:3FTCA	40.1	ng/L	40.00	ND	100	73-134	2.80	25.7	01/21/2026	
8:2FTS	41.3	ng/L	38.40	ND	108	76-135	7.02	26.3	01/21/2026	
9Cl-PF3ONS	38.8	ng/L	37.40	ND	104	76-136	3.89	23.9	01/21/2026	
ADONA	40.2	ng/L	37.80	ND	106	79-139	3.95	23	01/21/2026	
HFPO-DA	42.3	ng/L	40.00	ND	106	71-141	3.66	28.8	01/21/2026	
NEtFOSAA	42.6	ng/L	40.00	ND	107	77-135	0.943	24.6	01/21/2026	
NMeFOSAA	42.2	ng/L	40.00	ND	106	77-133	2.83	24.9	01/21/2026	
PFBA	44.1	ng/L	40.00	1.44	107	70-141	0.455	21.9	01/21/2026	
PFBS	37.2	ng/L	35.40	ND	105	79-137	2.81	25	01/21/2026	
PFBSA	42.4	ng/L	40.00	0.380	105	75-133	0.517	23.4	01/21/2026	
PFDA	42.9	ng/L	40.00	ND	107	77-135	4.42	26.3	01/21/2026	
PFDoDA	46.4	ng/L	40.00	ND	116	79-134	3.73	23.9	01/21/2026	

Organics-PFAS Isotope Dilution - Quality Control

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Analyzed	Qualifier
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	----------	-----------

Batch B6A1420 - Method: 3512

Prepared: 01/14/2026

Matrix Spike Dup (B6A1420-MSD1)

Source: 2601050-01

PFDS	36.8	ng/L	38.60	ND	95.4	68-132	3.52	28.4	01/21/2026	
PFECHS	37.1	ng/L	36.96	ND	100	62-153	6.96	23.1	01/21/2026	
PFHpA	41.9	ng/L	40.00	ND	105	77-134	0.571	24	01/21/2026	
PFHpS	39.8	ng/L	38.20	ND	104	75-134	0.604	25.7	01/21/2026	
PFHxA	42.9	ng/L	40.00	1.26	104	74-142	1.76	23	01/21/2026	
PFHxS	37.1	ng/L	36.50	ND	102	72-141	0.805	23.6	01/21/2026	
PFHxSA	41.4	ng/L	40.00	ND	103	76-133	2.11	24.1	01/21/2026	
PFNA	41.7	ng/L	40.00	ND	104	77-134	4.18	24	01/21/2026	
PFNS	37.7	ng/L	38.40	ND	98.2	76-134	6.47	24.1	01/21/2026	
PFOA	41.4	ng/L	40.00	1.18	101	71-141	1.15	22.5	01/21/2026	
PFOS	39.3	ng/L	37.04	1.08	103	71-144	5.68	23.4	01/21/2026	
PFOSA	40.6	ng/L	40.00	ND	101	78-131	1.81	23.4	01/21/2026	
PFPeA	43.2	ng/L	40.00	ND	108	72-142	0.697	22.9	01/21/2026	
PFPeS	37.7	ng/L	37.60	ND	100	70-139	2.77	24.4	01/21/2026	
PFTeDA	47.0	ng/L	40.00	ND	117	78-133	4.40	23.6	01/21/2026	
PFTrDA	45.8	ng/L	40.00	ND	115	66-136	5.52	29	01/21/2026	
PFUnDA	42.3	ng/L	40.00	ND	106	79-135	3.47	25.1	01/21/2026	
<i>Isotope Dilution Analog: 13C2-4:2FTS</i>	<i>161</i>	<i>ng/L</i>	<i>150.1</i>		<i>107</i>	<i>47-213</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C2-6:2FTS</i>	<i>132</i>	<i>ng/L</i>	<i>152.2</i>		<i>86.7</i>	<i>10-169</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C2-8:2FTS</i>	<i>158</i>	<i>ng/L</i>	<i>153.6</i>		<i>103</i>	<i>15-176</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C2-PFDoDA</i>	<i>148</i>	<i>ng/L</i>	<i>160.0</i>		<i>92.5</i>	<i>51-139</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C2-PFDoDA+13C2-PFTeDA</i>	<i>136</i>	<i>ng/L</i>	<i>160.0</i>		<i>85.3</i>	<i>38-128</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C2-PFTeDA</i>	<i>116</i>	<i>ng/L</i>	<i>160.0</i>		<i>72.4</i>	<i>10-123</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C3-HFPODA</i>	<i>340</i>	<i>ng/L</i>	<i>320.0</i>		<i>106</i>	<i>67-127</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C3-PFBS</i>	<i>146</i>	<i>ng/L</i>	<i>149.1</i>		<i>98.1</i>	<i>67-115</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C3-PFHxS</i>	<i>157</i>	<i>ng/L</i>	<i>151.7</i>		<i>104</i>	<i>87-116</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C4-PFBA</i>	<i>163</i>	<i>ng/L</i>	<i>160.0</i>		<i>102</i>	<i>72-121</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C4-PFHpA</i>	<i>168</i>	<i>ng/L</i>	<i>160.0</i>		<i>105</i>	<i>87-117</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C5-PFHxA</i>	<i>170</i>	<i>ng/L</i>	<i>160.0</i>		<i>106</i>	<i>85-116</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C5-PFPeA</i>	<i>165</i>	<i>ng/L</i>	<i>160.0</i>		<i>103</i>	<i>83-116</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C6-PFDA</i>	<i>179</i>	<i>ng/L</i>	<i>160.0</i>		<i>112</i>	<i>72-123</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C7-PFUnDA</i>	<i>166</i>	<i>ng/L</i>	<i>160.0</i>		<i>104</i>	<i>73-121</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	<i>174</i>	<i>ng/L</i>	<i>160.0</i>		<i>109</i>	<i>82-117</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C8-PFOS</i>	<i>157</i>	<i>ng/L</i>	<i>153.4</i>		<i>103</i>	<i>84-114</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C8-PFOA</i>	<i>166</i>	<i>ng/L</i>	<i>160.0</i>		<i>103</i>	<i>84-117</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: 13C9-PFNA</i>	<i>174</i>	<i>ng/L</i>	<i>160.0</i>		<i>109</i>	<i>83-123</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: d3-N-MeFOSAA</i>	<i>172</i>	<i>ng/L</i>	<i>160.0</i>		<i>107</i>	<i>76-130</i>			<i>01/21/2026</i>	
<i>Isotope Dilution Analog: d5-N-EtFOSAA</i>	<i>166</i>	<i>ng/L</i>	<i>160.0</i>		<i>104</i>	<i>74-136</i>			<i>01/21/2026</i>	